



Juniper Batch #: 20JA0246.01 A-B Composite
 Intake Date: 1/31/2020

Pesticide Analysis (Oregon Compliance Standard OAR 333-008-1190)

ANALYSIS DATE: 2/3/2020			Instrument: LC/MS/MS			Method: AOAC 2007.1-modified		
Pesticide	Result (ppm)	Action Level / LOQ (ppm)	Pesticide	Result (ppm)	Action Level / LOQ (ppm)			
Abamectin	<LOQ	0.5 / 0.25	Imazalil	<LOQ	0.2 / 0.10			
Acephate	<LOQ	0.4 / 0.20	Imidacloprid	<LOQ	0.4 / 0.20			
Acequinocyl	<LOQ	2.0 / 1.00	Kresoxim-methyl	<LOQ	0.4 / 0.20			
Acetamiprid	<LOQ	0.2 / 0.10	Malathion	<LOQ	0.2 / 0.10			
Aldicarb	<LOQ	0.4 / 0.20	Metalaxyl	<LOQ	0.2 / 0.10			
Azoxystrobin	<LOQ	0.2 / 0.10	Methiocarb	<LOQ	0.2 / 0.10			
Bifenazate	<LOQ	0.2 / 0.10	Methomyl	<LOQ	0.4 / 0.20			
Bifenthrin	<LOQ	0.2 / 0.10	Methyl Parathion	<LOQ	0.2 / 0.10			
Boscalid	<LOQ	0.4 / 0.20	MGK-264	<LOQ	0.2 / 0.10			
Carbaryl	<LOQ	0.2 / 0.10	Myclobutanil	<LOQ	0.2 / 0.10			
Carbofuran	<LOQ	0.2 / 0.10	Naled	<LOQ	0.5 / 0.25			
Chlorantraniliprole	<LOQ	0.2 / 0.10	Oxamyl	<LOQ	1.0 / 0.50			
Chlorfenapyr	<LOQ	1.0 / 0.50	Paclobutrazol	<LOQ	0.4 / 0.20			
Chlorpyrifos	<LOQ	0.2 / 0.10	Permethrins	<LOQ	0.2 / 0.10			
Clofentezine	<LOQ	0.2 / 0.10	Phosmet	<LOQ	0.2 / 0.10			
Cyfluthrin	<LOQ	1.0 / 0.50	Piperonyl butoxide	<LOQ	2.0 / 1.00			
Cypermethrin	<LOQ	1.0 / 0.50	Prallethrin	<LOQ	0.2 / 0.10			
Daminozide	<LOQ	1.0 / 0.50	Propiconazole	<LOQ	0.4 / 0.20			
DDVP (Dichlorvos)	<LOQ	1.0 / 0.50	Propoxur	<LOQ	0.2 / 0.10			
Diazinon	<LOQ	0.2 / 0.10	Pyrethrins	<LOQ	1.0 / 0.50			
Dimethoate	<LOQ	0.2 / 0.10	Pyridaben	<LOQ	0.2 / 0.10			
Ethoprophos	<LOQ	0.2 / 0.10	Spinosad	<LOQ	0.2 / 0.10			
Etofenprox	<LOQ	0.4 / 0.20	Spiromesifen	<LOQ	0.2 / 0.10			
Etoxazole	<LOQ	0.2 / 0.10	Spirotetramat	<LOQ	0.2 / 0.10			
Fenoxycarb	<LOQ	0.2 / 0.10	Spiroxamine	<LOQ	0.4 / 0.20			
Fenpyroximate	<LOQ	0.4 / 0.20	Tebuconazole	<LOQ	0.4 / 0.20			
Fipronil	<LOQ	0.4 / 0.20	Thiacloprid	<LOQ	0.2 / 0.10			
Flonicamid	<LOQ	1.0 / 0.50	Thiamethoxam	<LOQ	0.2 / 0.10			
Fludioxonil	<LOQ	0.4 / 0.20	Trifloxystrobin	<LOQ	0.2 / 0.10			
Hexythiazox	<LOQ	1.0 / 0.50						
Pesticide Screen	Pass							

LOQ= Limit of Quantification

Microbiological Contaminants (Oregon Compliance Standard OAR 333-007-0390)

ANALYSIS DATE: Not Tested

Microbiological screening	Colony count	CFU/g	Results:
Total coliforms	Not tested	Not tested	N/A
<i>Escherichia coli</i> (E. coli)	Not tested	Not tested	N/A